

Henry Ford Health System

## Henry Ford Health System Scholarly Commons

---

Dermatology Articles

Dermatology

---

4-14-2020

### The importance of the minimum dosage necessary for UVC decontamination of N95 respirators during the COVID-19 pandemic

Shanthi Narla

*Henry Ford Health System, snarla1@hfhs.org*

Alexis B. Lyons

*Henry Ford Health System, ALyons5@hfhs.org*

Indermeet Kohli

*Henry Ford Health System, ikohli1@hfhs.org*

Angeli E. Torres

*Henry Ford Health System, atorres5@hfhs.org*

Angie Parks-Miller

*Henry Ford Health System, AMILLER5@hfhs.org*

*See next page for additional authors*

Follow this and additional works at: [https://scholarlycommons.henryford.com/dermatology\\_articles](https://scholarlycommons.henryford.com/dermatology_articles)

---

#### Recommended Citation

Narla S, Lyons AB, Kohli I, Torres AE, Parks-Miller A, Ozog DM, Hamzavi IH, and Lim HW. The Importance of the Minimum Dosage Necessary for UVC Decontamination of N95 Respirators during the COVID-19 Pandemic. *Photodermatol Photoimmunol Photomed* 2020.

This Article is brought to you for free and open access by the Dermatology at Henry Ford Health System Scholarly Commons. It has been accepted for inclusion in Dermatology Articles by an authorized administrator of Henry Ford Health System Scholarly Commons.

---

**Authors**

Shanthi Narla, Alexis B. Lyons, Indermeet Kohli, Angeli E. Torres, Angie Parks-Miller, David M. Ozog, Iltefat H. Hamzavi, and Henry W. Lim



# The importance of the minimum dosage necessary for UVC decontamination of N95 respirators during the COVID-19 pandemic

## To the Editor,

The World Health Organization (WHO) recently released a press report highlighting the severe shortage of personal protective equipment (PPE) that is endangering healthcare workers worldwide during the COVID-19 pandemic.<sup>1</sup> To meet this urgent need, healthcare institutions across the world have begun to utilize the germicidal properties of ultraviolet C (UVC) to decontaminate N95 respirators so that they can be reused.<sup>2</sup> It is clearly crucial that the dose of UVC delivered is sufficient to kill any viable SARS-CoV-2, the causative virus of the COVID-19 pandemic that may be present on the respirators.

Currently, there is a significant amount of variability in the UVC dosages used to decontaminate N95 respirators.<sup>3,4</sup> While no published data are available on the UVC dose needed to decontaminate respirators from SARS-CoV-2, other coronaviruses such as severe acute respiratory syndrome coronavirus (SARS-CoV) and Middle East respiratory syndrome coronavirus (MERS-CoV)<sup>5</sup> could potentially serve as surrogate markers. Heimbuch et al evaluated the doses of UVC needed to decontaminate circular coupons prepared from 3M 1870 N95 filtering facepiece respirators (FFRs) using a variety of viruses including influenza A (H1N1), avian influenza A virus (H5N1), influenza A (H7N9) A/Anhui/1/2013, influenza A (H7N9) A/Shanghai/1/2013, MERS-CoV, and SARS-CoV. There was no detectable viable virus for all six strains after UVC decontamination using 1 J/cm<sup>2</sup>.<sup>6</sup> This dose is now reflected in the Centers for Disease Control and Prevention (CDC)<sup>7</sup> and N95Decon's websites.<sup>8</sup>

However, it is important to note that in the above study, the viruses were only tested on circular coupons from one type of N95 respirator. As Mills et al<sup>9</sup> and Heimbuch et al<sup>6</sup> further illustrated in their studies, the 1 J/cm<sup>2</sup> dose may not be adequate to kill the tested viruses depending on the material/type of respirator used. Mills et al found that only facepieces on 12 of 15 models and straps on 7 of 15 models showed a significant ( $\geq 3$  log) reduction of H1N1 influenza viability. Similarly, Heimbuch et al found that only facepieces on 11 of 15 models and straps on 4 of 15 models showed a significant ( $\geq 3$  log) reduction of H1N1 influenza viability. In addition, there were differences in many cases between whether mucin or sebum was used as an artificial soiling agent, highlighting the importance of making sure that no shadowing (when materials such as cosmetics or sunscreens are deposited on the respirators)

occurs because this can block UVC penetration and hinder UVC decontamination.<sup>6,8</sup>

During this crisis, there is a myriad of UVC devices being discussed for possible use in health care settings which have variable output and design. There is the urgent need to raise awareness that at least 1 J/cm<sup>2</sup> is used to all surfaces for UVC decontamination methods to ensure that we are not endangering our healthcare workers. It should also be emphasized that there are significant limitations to UVC decontamination methods due to the variety of respirators used in healthcare facilities. Consequently, this process should only be considered as a risk mitigation effort during severe shortage of N95 respirators but is one of the most effective and best studied options available to front-line personnel.

## KEYWORDS

coronaviruses, COVID-19, dosage, filtering facepiece respirators, N95, pandemic, reuse, risk mitigation, SARS-CoV-2, ultraviolet C

## CONFLICTS OF INTEREST

DO is an investigator for Biofrontera. ABL and SN are sub-investigators for Biofrontera. IK, APM, and AET have no relevant disclosures. HWL is an investigator for LITE study which is funded by PCORI, and home phototherapy machines are provided by Daavlin, and has participated as a speaker in general educational session for Ra Medical System. IHH is an investigator for LITE study which is funded by PCORI, and home phototherapy machines are provided by Daavlin.

Shanthi Narla<sup>1</sup>   
Alexis B. Lyons<sup>1</sup>   
Indermeet Kohli<sup>1,2</sup>   
Angeli E. Torres<sup>1</sup>   
Angela Parks-Miller<sup>1</sup>   
David M. Ozog<sup>1</sup>   
Iltefat H. Hamzavi<sup>1</sup>   
Henry W. Lim<sup>1</sup>

<sup>1</sup>Photomedicine and Photobiology Unit, Department of Dermatology, Henry Ford Hospital, Detroit, MI, USA

<sup>2</sup>Department of Physics & Astronomy, Wayne State University, Detroit, MI, USA

## Correspondence

Henry W. Lim, MD, Department of Dermatology, Henry Ford  
Medical Center New Center One, 3031 W. Grand Blvd Suite  
800, Detroit, MI 48202, USA  
Email: hlim1@hfhs.org

## ORCID

Shanthi Narla  <https://orcid.org/0000-0003-0228-3262>  
Alexis B. Lyons  <https://orcid.org/0000-0003-0799-5680>  
Indermeet Kohli  <https://orcid.org/0000-0002-2316-6375>  
Angeli E. Torres  <https://orcid.org/0000-0002-9551-2118>  
Angela Parks-Miller  <https://orcid.org/0000-0001-5605-5842>  
David M. Ozog  <https://orcid.org/0000-0001-8479-6736>  
Iltefat H. Hamzavi  <https://orcid.org/0000-0002-3137-5601>  
Henry W. Lim  <https://orcid.org/0000-0002-1576-1115>

## REFERENCES

1. Organization WH. Shortage of personal protective equipment endangering health workers worldwide. 2020; <https://www.who.int/news-room/detail/03-03-2020-shortage-of-personal-protective-equipment-endangering-health-workers-worldwide>. Accessed 9 April 2020.
2. Hamzavi IH, Lyons AB, Kohli I, et al. Ultraviolet germicidal irradiation: possible method for respirator disinfection to facilitate reuse during COVID-19 pandemic. *J Am Acad Dermatol*. 2020. Epub ahead of print. <https://doi.org/10.1016/j.jaad.2020.03.085>.
3. Lowe JJ, Paladino KD, Farke JD, Boulter K, Cawcutt K, Emodi M, Gibbs S, Hankins R, Hinkle L, Micheels T, Schwedhelm S, Vasa A, Wadman M, Watson S, Rupp ME. N95 Filtering Facepiece Respirator Ultraviolet Germicidal Irradiation (UVGI) Process for Decontamination and Reuse 2020; <https://www.nebraskamed.com/sites/default/files/documents/covid-19/n-95-decon-process.pdf>. Accessed 9 April 2020.
4. Crozier KJ, Card D, Dhawan A, Dinh M, Dolson E, Farrokhan N, Gopalakrishnan V, Ho E, King ES, Krishnan N, Kuzmin G, Maltas J, Pelesko J, Scarborough JA, Scott JG, Sedor G, Weaver DT. UV Sterilization of Personal Protective Equipment with Idle Laboratory Biosafety Cabinets During the COVID-19 Pandemic 2020; <https://www.medrxiv.org/content/10.1101/2020.03.25.20043489v1.full.pdf>. Accessed 9 April 2020.
5. Bedell K, Buchaklian AH, Perlman S. Efficacy of an automated multiple emitter whole-room ultraviolet-C disinfection system against Coronaviruses MHV and MERS-CoV. *Infect Control Hosp Epidemiol*. 2016;37(5):598-599.
6. Heimbuch BK, Harnish D. Research to Mitigate a Shortage of Respiratory Protection Devices During Public Health Emergencies. 2020; <https://www.ara.com/news/ara-research-mitigate-shortage-respiratory-protection-devices-during-public-health-emergencies>. Accessed 9 April 2020.
7. CfDCa P. Decontamination and Reuse of Filtering Facepiece Respirators. 2020; <https://www.cdc.gov/coronavirus/2019-ncov/hcp/ppe-strategy/decontamination-reuse-respirators.html>. Accessed 9 April 2020.
8. N95DECON. Technical Report for UV-C-Based N95 Residue Risk Management. 2020; [https://static1.squarespace.com/static/5e8126f89327941b9453eeef/t/5e8541760211467623b1e4e4/1585791351199/200401\\_N95DECON\\_UV\\_technicalreport\\_v1.2\\_final.pdf](https://static1.squarespace.com/static/5e8126f89327941b9453eeef/t/5e8541760211467623b1e4e4/1585791351199/200401_N95DECON_UV_technicalreport_v1.2_final.pdf). Accessed 9 April 2020.
9. Mills D, Harnish DA, Lawrence C, Sandoval-Powers M, Heimbuch BK. Ultraviolet germicidal irradiation of influenza-contaminated N95 filtering facepiece respirators. *Am J Infect Cont*. 2018;46(7):e49-e55.